

CERTIFICATE OF ANALYSIS

Client: Asami Co., Ltd.
Crystal Clear Waters
1-33-11 Higashi-Izumi, Komae-shi, Tokyo 201-0014, Japan

Sample name: Water bottle A.
Production date 2011.3.23 Expiry date 2011.9.18 Samukawa Plant

Received date: March 25, 2011

This is to certify that the following result(s) have been obtained from our analysis on the above-mentioned sample(s) submitted by the client.

Test Result(s)

Test Item	Result	QL	N	M
Radioactive nuclide	1	
Cesium-137	Not detected	1.8 Bq/kg	2	
Cesium-134	Not detected	2.1 Bq/kg	2	
Iodine-131	Not detected	20 Bq/kg	2	

QL: Quantitation limit N: Notes M: Method

Notes

- 1: γ -ray spectrometer (germanium semiconductor detector).
2: Measuring period, 1000 seconds.



Noriko Imaizumi
Principal Investigator



Date